

LORAZEPAM HPLC-UV QC information

11/10/11 Run date: (Initial Sample Run Date)

Sample	Result (mg/mL)	% Recovery
QC-1	1.65	83
QC-2	1.68	84

11/15/11 Run Date: (QC Pipet Test)

Sample	Result (mg/mL)	% Recovery	Pipet Type
QC-RP1	1.79	90	Reverse Pipetting
QC-RP2	1.74	87	Reverse Pipetting
QC-RP3	1.78	89	Reverse Pipetting
QC-RW1	1.81	91	Regular pipet with tip wash
QC-RW2	1.75	88	Regular pipet with tip wash
QC-RW3	1.85	93	Regular pipet with tip wash

Average Recovery of Reverse Pipetting = 89%

Average Recovery of regular pipetting with tip wash = 91%